



## LABORATORY REPORT

FAWC  
1735 St Julian Pl Ste 101  
Columbus, SC 29204  
  
Phone: 839-213-2665

Account #: E73555  
Sample Name: KLOW  
Lot #: KL6926  
Storage Conditions: Refrigerated

Date Received: 6/10/2026

Submission #: ETX-260609-0112



### **Microbiology Tests:**

#### **ScanRDI Sterility Test**

**Method Suitability:** No Method Suitability on file

**Result:** Pass

Date Started: 6/10/2026

Date Completed: 6/11/2026

#### **USP <85> Bacterial Endotoxin Test**

**Numerical Value :** <2 EU/mL

**Endotoxin Limit:** 875 EU/mL

**Result:** Pass

Date Started: 6/12/2026

Date Completed: 6/12/2026

Respectfully submitted,  
EAGLE ANALYTICAL SERVICES INC.

Alison Eng, Quality Control Manager

**Results apply only to the tested sample(s) as received.**